

Ludger

## Certificate of Analysis

### LudgerPure™ 2AB Labeled M3N2Glycan

Cat. #: CAB-M3N2-01

Lot #: A81H-05

Size : approx. 100 pmol

**Purity:** > 90% pure as assessed by HPLC (see Fig 1)




Figure 1 : LudgerSepN1 HPLC profile of 2AB labeled M3N2 glycan (Cat. #: CAB-M3N2-01, Lot No. A81H-05).

#### HPLC Running Conditions

Column: LudgerSep N1 Amide (Cat. #: LS-N1-4.6x250)      Flow: 0.4 – 1.0 ml/min      Temperature: 40 °C

Solvent A: 100 % acetonitrile

Solvent B: 50 mM ammonium formate pH 4.4

Gradient: 0-152 min – 20-58 % B

152-155 min - 58-100 % B

155-157 min - 100 % B

157-162 min - 100% B (flow 1 ml/min)

162-163 min – 100-20 % B

163-177 min – 20 % B

177-178 min – flow 1 ml/min–0.4 ml/min

178-180 min – 20 % B (flow 0.4 ml/min)

Detector: Jasco FP920

Excitation wavelength: 360 nm

Emission wavelength: 425 nm