

Ludger

## Certificate of Analysis


### LudgerPure™ 2AB Labeled NGA2F Glycan

Cat. #: CAB-NGA2F-01

Lot #: A51C-01

Size: approx. 100 pmol

**Purity:** > 90% pure as assessed by HPLC (see Fig 1)



**Figure 1 : LudgerSepN1 HPLC profile of 2AB labeled NGA2F glycan (Cat. #: CAB-NGA2F-01, Lot No. A51C-01).**

#### HPLC Running Conditions

Column: LudgerSep N1 Amide (Cat. #: LS-N1-4.6x250)

Flow: 0.4 ml/min

Temperature: 40 °C

Solvent A: 100 % acetonitrile

Solvent B: 250 mM ammonium formate pH 4.4

Gradient: 0-75 min - 35-65 % B

75-80 min - 65-100 % B

80-83 min - 100 % B

83-85 min - 100-35 % B

85-115 min - 35 % B

Detector: Jasco FP920

Excitation wavelength: 360 nm

Emission wavelength: 425 nm