

Certificate of Analysis

LudgerSep N2 Amide HPLC Glycan Analysis Column

Cat. # : LS-N2-4.6x150

Serial # : B15R-12

Size : 4.6 x150mm

This column conforms to the specifications given in Ludger document # LS-N2-GUIDE.

Packed Solvent 85 % acetonitrile (aq)

Test Analysis

Flow Rate 1.0 ml / min

Eluent Isocratic 85 % acetonitrile (aq)

Pressure 4.9 MPa


Temperature 40 °C

Sample Uracil 35mg/L

Sample Volume 10 µl

Detection UV 254nm

Inspection Chromatogram



Theoretical Plates = 20622

Asymmetry Factor = 1.07