



**Product Guide for LudgerSep™ N2 High Resolution
Amide HPLC Columns
for Glycan Analysis**

(Ludger Product Codes: LS-N2-4.6x150 and LS-N2-2.0x150)

Ludger Document # LS-N2-Guide-v4.0

Ludger Ltd

Culham Science Centre
Oxford OX14 3EB
United Kingdom

Tel: +44 1865 408 554

Fax: +44 870 163 4620

Email: info@ludger.com

www.ludger.com

Contents

	Page
Contents.....	2
Specifications for LudgerSep™ N2 Columns.....	3
HPLC System Requirements	5
Installation of the Column	5
Preconditioning of the Newly Installed Column.....	6
Column Cleaning and Storage.....	6
Light Contamination.....	6
Cleaning Medium Contamination	6
Cleaning Heavy Contamination	6
Long Term Storage.....	7
Sample Preparation	7
Fluorescent Labeling	7
Filtering Samples	8
Sialylated Glycans	8
Operating Schedule and System Suitability Test for 4.6 x 150 mm Column	8
Glycan Analysis with the 4.6 x 150 mm LS-N2 Amide HPLC Column.....	11
Solvents	11
Gradient # LSN2-10m-Start-35% : N-Glycan Analysis Startup	11
Gradient # LSN2-30m-Nlink-35% : N-Glycan Analysis Gradient	11
Gradient # LSN2-10m-Start-20% : O-Glycan and N-Glycan Analysis Startup.....	12
Gradient # LS-N2-60m-ONlink-20% : O-Glycan and N-Glycan Analysis Gradient.....	12
Operating Schedule and System Suitability Test for two 2.0 x 150 mm Columns	13
Glycan Analysis with two 2 x 150 mm LS-N2 Amide HPLC Columns.....	15
Solvents	15
Gradient # LSN2x2-5m-Start-30% : N-Glycan Analysis Startup	15
Gradient # LSN2x2-30m-Nlink-30% : N-Glycan Analysis Gradient.....	15
Warranties and Liabilities.....	16
Document Revision Number	16

Specifications for LudgerSep™ N2 Columns

Application Analysis and purification by HPLC of LudgerTag™ fluorophore and UV-chromophore labeled glycans. The 2mm diameter column is recommended for LC-MS applications.

Description LudgerSep™ N2 HPLC columns contain particles with a polymeric amide coating optimized for high resolution chromatography of complex glycan mixtures.

Particles 3 µm particle size with 80 angstrom pores and polymeric amide coating.

Column Size	Cat #	Description	Dimensions
	LS-N2-4.6x150	LudgerSep N2 HPLC Column	4.6 x 150 mm
	LS-N2-2.0x150	LudgerSep N2 HPLC Column	2.0 x 150 mm

Flow Rates Typical flow rates = 0.4 - 1.0 ml/min for 4.6mm column
= 0.15 - 0.22ml/min for 2.0mm column

Column Pressure Maximum pressure = 2250 psi (150 kg/cm²) for 4.6mm column
= 2900 psi (195 kg/cm²) for 2.0mm column

pH Range 2.0 - 7.5

Temperature Typical operating temperature = 30 °C.
Temperature range = 10 - 40 °C.

Solvents Typical solvent systems for glycan analysis include gradients of acetonitrile and aqueous buffers containing ammonium formate, pH 4.4. The recommended buffer is 50 mM ammonium formate pH 4.4 which is available in 50 ml aliquots for making up 2 liters: (Ludger catalog # : LS-N-BUFFX40)

Shipping Solvent 75% acetonitrile - 25% water

Storage For long-term storage, the column should be washed with a gradient going from the operational buffer to 100% water then to the storage solvent of >50% acetonitrile in water (v/v).

Column Protection Filter all solvents to 0.2 µm and degas using either helium sparging or vacuum degassing.
Filter all samples using a 0.2 µm filter membrane before loading onto the column.
Install an in-line filter between the sample injector and the column.

Amount of Sample The maximum amount of glycan sample that can be loaded on the column depends on the number and type of glycan components as well as the nature of any non-glycan material. The typical range for successful analytical runs is 10fmol - 1 nmol per sample peak and up to 200 nmol of total glycans.

Suitable Samples Suitable samples include glycans labeled with the following LudgerTag™ labels : 2-AA (2-aminobenzoic acid), 2-AB (2-aminobenzamide), AA-Ac (3-(acetylamino)-6-aminoacridine

Sample Filter samples to 0.2 µm then dry using a centrifugal evaporator.

Preparation For the 4.6mm column inject the sample in up to 100 µl of the starting buffer (i.e. the solvent mixture used at the very start of the HPLC gradient).

For the 2mm column inject the sample in up to 25 µl of the starting buffer (i.e. the solvent mixture used at the very start of the HPLC gradient)

Sample Detection Either fluorescence, mass spectrometry or UV-absorbance depending on the dye used (see the appropriate LudgerTag™ instruction guide).

Handling: Ensure that any glass, plasticware or solvents used are free of glycosidases and environmental carbohydrates. Use powder-free gloves for all sample handling procedures and avoid contamination with environmental carbohydrate.

Safety: Please read the Material Safety Data Sheets (MSDS's) for all chemicals used. All processes involving labeling reagents should be performed using appropriate personal safety protection - eyeglasses, chemically resistant gloves (e.g. nitrile), etc. - and where appropriate in a laboratory fume cupboard

For research use only. Not for human or drug use

HPLC System Requirements

LudgerSep™ N2 columns can be used with an HPLC system capable of delivering accurate gradients at a flow rate of 0.3 to 1.0 ml/min for the 4.6 mm diameter column, or 50 to 300 µl/min for the 2mm diameter column. In general, systems which mix eluants at high pressure (after the pump head) have lower dead volumes and supply more accurate gradients that are appropriate at the flow rate needed for LudgerSep™ columns.

For the 4.6mm column inject the sample in up to 100 µl of the starting buffer (i.e. the solvent mixture used at the very start of the HPLC gradient). For the 2mm column inject the sample in up to 25 µl of the starting buffer

A fluorescence detector is required with the following detection wavelengths:

Fluorescence Label		λ_{ex} (nm)	λ_{em} (nm)
2-AB	[2-aminobenzamide]	330	420
2-AA	[2-aminobenzoic acid]	330	420
AA-Ac	[3-(acetylamino)-6-aminoacridine]	442	525

For optimal detection, use wide slit widths (e.g. 10 – 20 nm). Sub-picomole levels of 2-AB or 2-AA labelled glycans can be detected with good signal-to-noise (depending on the sensitivity of the detector used).

To improve repeatability and intermediate precisions for glycan analyses use a column temperature controller. Good results can be obtained with a column temperature of 35°C.

Installation of the Column

During column installation we recommend that :

- You should connect the LudgerSep™ N2 column to your HPLC system using standard 1/16" OD tubing and 10-32 (1/16") fittings in either stainless steel or PEEK (polyetheretherketone). Finger-tight PEEK fittings and tubing (0.17 mm / 0.007" ID) are recommended for ease of connection and to minimise damage to the column threads. Flow direction is indicated by an arrow on the column label.
- Keep the lengths of tubing between the injector to column and column to detector as short as possible to minimise dispersion effects.
- Install an in-line 0.2 µm filter with minimal dead volume between the injector and the head of the

LudgerSep™ N2 column to prevent damage to the column by particles.

- Where two 150 x 2mm columns are joined together use as short a length of tubing as possible.
- Before analysing any samples, condition your newly installed column as follows

Preconditioning of the Newly Installed Column

Prepare your newly installed column for glycan analysis as follows:

- Precondition the column by running two complete LSN2-30m-Nlink-35% gradient cycles (see below) without any sample injected.
- Check the column performance by running a fluorescently labeled glucose homopolymer (GHP) as a system suitability standard using a LSN2-30m-Nlink-35% gradient for the 4.6 mm column or a LSN2x2-60m-Nlink-30% gradient for two 2mm columns (see section on suitability suitability below).

Column Cleaning and Storage

After heavy use, your LudgerSep™ N2 column may become contaminated with strongly adsorbed sample constituents that will lead to a loss in column performance.

Light Contamination

- 1 Clean the column following the protocol in the 'Column Preconditioning' section above.
- 2 Equilibrate in the starting buffer for the LSN2-30m-Nlink-35% (4.6 mm column) or LSN2x2-60m-Nlink-30% (2 mm columns) gradient
- 3 Do a blank run without injecting a sample to check the baseline.

Cleaning Medium Contamination

- 1 Run through one LSN2-30m-Nlink-35% (4.6 mm column) or LSN2x2-60m-Nlink-30% (2mm columns) gradient.
- 2 Wash with 50% (v/v) 50 mM ammonium formate pH 4.4 (cat. No. LS-N-BUFFX40) 50% (v/v) acetonitrile at 0.1 ml/min overnight
- 3 Run through one LSN2-30m-Nlink-35% (4.6 mm column) or LSN2x2-60m-Nlink-30% (2 mm columns) gradient
- 4 Run a system suitability check

Cleaning Heavy Contamination

Wash the column with the following:

- 1 Water [4 h or more at 0.1 ml/min] - *to remove very polar solutes from the bonded phase*
- 2 45% acetonitrile (aq) [4 h or more at 0.1 ml/min] - *to desorb hydrophobic compounds*
- 3 0.1% triethylamine in 80% acetonitrile [4 h or more at 0.1 ml/min] - *to remove desorbed basic compounds*
- 4 50 mM ammonium formate pH 4.4 / acetonitrile (1:1 v/v) [4 h or more at 0.1 ml/min] - *to remove ionic compounds*
- 5 Run through one LSN2-30m-Nlink-35% gradient (4.6 mm column) or LSN2x2-60m-Nlink-30% (2 mm columns)
- 6 Run a system suitability check

Long Term Storage

For long-term storage, the column should be washed with a gradient going from the operational buffer to 100% water then to the storage solvent of >50% acetonitrile in water (v/v).

Sample Preparation

LudgerSep™ N2 columns should be used for analysis of purified glycans (fluorescently labeled or unlabeled). Samples must be free of particulates. The sample should be dissolved in water **before** addition of acetonitrile to make up to the same composition as the start of the gradient (e.g. dissolve sample in 35 µl water, then add 65 µl acetonitrile before injection to run with the N-glycan analysis gradient method: LSN2-30m-Nlink-35%).

Fluorescent Labeling

Fluorescent glycans can be prepared by derivatizing pure glycans using reductive amination with a LudgerTag™ label and purification on a LudgerClean™ cartridge. The following are commonly used labeling and purification systems for biopharmaceutical glycosylation analysis:

Labeling System

2-AB (LudgerTag™ kit # LT-KAB-A2)

2-AA (LudgerTag™ kit # LT-KAA-A2)

AA-Ac (LudgerTag™ kit # LT-KAAAC-A2)

Post-Labeling Purification System

LudgerClean™ S-Cartridge (Cat # LC-S-A6)

LudgerClean™ S-Cartridge (Cat # LC-S-A6)

LudgerClean™ D1-Cartridge (Cat # LC-D1-A6)

Filtering Samples

Remove particulates from samples by filtering through a spin filter or syringe filter with 0.2 µm pore size membrane.

Sialylated Glycans

Sialylated glycans can become desialylated if exposed to acidic conditions and elevated temperatures. Avoid desialylation with such samples by

- a. minimizing exposure to acid (if possible, keep the pH between 5 - 8), and
- b. minimizing exposure to temperatures greater than 25°C.

Operating Schedule and System Suitability Test for 4.6 x 150 mm Column

- Two running methods are listed below: the 30 min run starting at 35% A is suitable for analysis of N-glycans; the 60 min run starting at 20% A is suitable for analysis of both O-glycans and N-glycans
- When starting the analysis bring the flow rate up slowly by running the startup method: LSN2-10m-Start-35% for N-glycan analysis or LSN2-10m-Start-20% for O-glycan and N-glycan analysis
- Follow this by running a complete gradient with no sample to condition the column with the running method: LSN2-30m-Nlink-35% : N-Glycan Analysis Gradient or LS-N2-60m-ONlink-20% (see below for gradient methods)
- Run one of the following system suitability standards using the LSN2-30m-Nlink-35% or LS-N2-60m-ONlink-20% gradient:

Product #	Description
CAB-GHP-30	2-AB GHP (2-AB labelled glucose homopolymer ladder)
CAA-GHP-30	2-AA GHP (2-AA labelled glucose homopolymer ladder)




Figure 1: Typical HPLC chromatogram from 2-AB GHP run with method LSN2-30m-Nlink-35% on a 4.6 x 150 mm LudgerSepN2 column.

Sample injected in 35%aqueous, 65% acetonitrile.




Figure 2: Typical HPLC chromatogram from 2-AB GHP run with method LSN2-60m-Nlink-20% on a 4.6 x 150 mm LudgerSepN2 column.

Sample injected in 20%aqueous, 80% acetonitrile.

- The GHP system suitability profile should be similar to that in figure 1 using the gradient LSN2-30m-Nlink-35%, or figure 2 if using the gradient LSN2-60m-Nlink-20%,.
- If all 12 peaks do not elute in this time, then this is likely to be due to a high dead volume in the HPLC system. Increase the run time to take this into account. (e.g. gradient of 35-47% A over 24 min)
- The peaks should be symmetrical, if there is tailing or the peaks are too broad then try cleaning the column (see Section: Column Cleaning and Storage) or change to a new column.
- Continue running fluorescent labelled GHP suitability standards approximately every 10 sample runs

- The GHP ladder can also be used to calibrate against day-to-day and system-to-system changes by expressing the elution time of peaks as glucose units (GU). The GU value is calculated by fitting a cubic spline distribution curve to the GHP ladder (GU 1-12 in figure 1), this curve is then used to allocate GU values from retention times. These GU values are very reproducible for neutral glycans (± 0.03), with more variation in sialylated glycans (± 0.3) and can be compared to database values [<http://glycobase.ucd.ie/cgi-bin/public/glycobase.cgi>]

Glycan Analysis with the 4.6 x 150 mm LS-N2 Amide HPLC Column

Solvents

The glycan analysis gradients in this guide are based on the following solvents:

- Solvent A** : 50 mM ammonium formate pH 4.4,
(50 ml of Ludger product # LS-N-BUFFX40 diluted to 2 litres with water)
- Solvent B** : Acetonitrile

Gradient # LSN2-10m-Start-35% : N-Glycan Analysis Startup

Use as a start up method for gradient for analysis of N-glycans which starts at 35% A.

Time (min)	% A	%B	4.6mm column Flow Rate (ml/min)
0	35	65	0.0
2	35	65	1.0
10	35	65	1.0

Gradient # LSN2-30m-Nlink-35% : N-Glycan Analysis Gradient

Use as a gradient for analysis of N-glycans with GU between 4 and 12.

Time (min)	% A	%B	4.6mm column Flow Rate (ml/min)
0	35	65	1.0
22	46	54	1.0
22.5	100	0	1.0
24.5	100	0	1.0
26	35	65	1.0
30	35	65	1.0

Gradient # LSN2-10m-Start-20% : O-Glycan and N-Glycan Analysis Startup

Use as a start up method for gradient for analysis of O-glycans and N-glycans which starts at 20% A.

Time (min)	% A	%B	4.6mm column Flow Rate (ml/min)
0	20	80	0.0
2	20	80	1.0
10	20	80	1.0

Gradient # LS-N2-60m-ONlink-20% : O-Glycan and N-Glycan Analysis Gradient

Use as a gradient for analysis of O-glycans and N-glycans with GU between 1 and 12.

Time (min)	% A	%B	4.6mm column Flow Rate (ml/min)
0	20	80	1.0
52	46	54	1.0
52.5	100	0	1.0
54.5	100	0	1.0
54	20	80	1.0
60	20	80	1.0

Operating Schedule and System Suitability Test for two 2.0 x 150 mm Columns

Two 2.0 x 150 mm columns can be joined together in series to obtain double the column length. This setup produces similar resolution to that obtained using one 4.6 x 150 mm column, with the advantage of cutting solvent use to a third of that required for the larger diameter column. A single column can be used but the resolution is not as high.

- The method for N-glycan analysis is listed below.
- When starting the analysis bring the flow rate up slowly by running the startup method: LSN2x2-5m-Start-30%
- Follow this by running a complete gradient with no sample to condition the column with the running method: LSN2x2-60m-Nlink-30%
- Run one of the following system suitability standards using the LSN2x2-60m-Nlink-30%

Product #	Description
CAB-GHP-30	2-AB GHP (2-AB labelled glucose homopolymer ladder)
CAA-GHP-30	2-AA GHP (2-AA labelled glucose homopolymer ladder)




Figure 3: Typical HPLC chromatogram from 2-AB GHP run with method LSN2-30m-Nlink-35% on two 2 x 150 mm LudgerSepN2 columns in series.

Sample injected in 30%aqueous, 70% acetonitrile.

- The GHP system suitability profile should be similar to that in figure 3 using the gradient LSN2x2-60m-Nlink-30%.
- If enough peaks do not elute in this time, then this is likely to be due to a high dead volume in the HPLC system. Increase the run time to take this into account.
- The peaks should be symmetrical, if there is tailing or the peaks are too broad then try cleaning the column (see Section: Column Cleaning and Storage) or change to a new column.
- Continue running fluorescent labelled GHP suitability standards approximately every 10 sample runs
- The GHP ladder can also be used to calibrate against day-to-day and system-to-system changes by expressing the elution time of peaks as glucose units (GU). The GU value is calculated by fitting a cubic spline distribution curve to the GHP ladder (GU 1-12 in figure 1), this curve is then used to allocate GU values from retention times. These GU values are very reproducible for neutral glycans (± 0.03), with more variation in sialylated glycans (± 0.3) and can be compared to database values [<http://glycobase.ucd.ie/cgi-bin/public/glycobase.cgi>]

Glycan Analysis with two 2 x 150 mm LS-N2 Amide HPLC Columns

Solvents

The glycan analysis gradients in this guide are based on the following solvents:

Solvent A : 50 mM ammonium formate pH 4.4,
(50 ml of Ludger product # LS-N-BUFFX40 diluted to 2 litres with water)

Solvent B : Acetonitrile

Gradient # LSN2x2-5m-Start-30% : N-Glycan Analysis Startup

Use as a start up method for gradient for analysis of N-glycans which starts at 30% A.

Time (min)	% A	%B	2 x 2mm column Flow Rate (ml/min)
0	30	70	0.0
2	30	70	0.18
5	30	70	0.18

Gradient # LSN2x2-30m-Nlink-30% : N-Glycan Analysis Gradient

Use as a gradient for analysis of N-glycans with GU between 4 and 12.

Time (min)	% A	%B	2 x 2mm column Flow Rate (ml/min)
0	30	70	0.18
46	53	47	0.18
46.5	100	0	0.18
49.5	100	0	0.18
51	30	70	0.18
60	30	70	0.18

Warranties and Liabilities

Ludger warrants that the above product conforms to the attached analytical documents. Should the product fail for reasons other than through misuse Ludger will, at its option, replace free of charge or refund the purchase price. This warranty is exclusive and Ludger makes no other warrants, expressed or implied, including any implied conditions or warranties of merchantability or fitness for any particular purpose.

Ludger shall not be liable for any incidental, consequential or contingent damages.

This product is intended for *in vitro* research only.

Document Revision Number

Document # LS-N2-Guide-v4.0