



## Certificate of Analysis

### LudgerSep UR2 HPLC Column for Glycan Analysis

Cat. #: LS-UR2-2.1x50

Batch #: B1CC-04

Size: 2.1 x 50 mm

This column conforms to the specifications given in Ludger document # LS-UR2-GUIDE.

**Packed Solvent** 50 % acetonitrile (aq)

#### Test Analysis

**Flow Rate** 500ul / min


**Eluent** Isocratic 50 % acetonitrile (aq)

**Pressure** 5800 psi

**Temperature** Ambient

**Sample Volume** 1  $\mu$ l

**Detection** UV at 254nm



Peak No	Peak Name	Retention time (minutes)	Tailing Factor (5%)
1	Theophyline	0.25	1.63
2	p-Nitroaniline	0.41	1.45
3	Methyl Benzoate	0.68	1.22
4	Phenetole	1.04	1.10
5	o-Xylene	1.57	1.06

Efficiency o-Xylene (n/m) = 195158

Tailing Factor (5%) o-Xylene = 1.06