

Certificate of Analysis

NeuGc Standard

Cat. #: CM-NEUGC-01

Batch: B7A4-01

Size : 1.04 ± 0.04 nmol

Expiry Date: 28 Nov 2022

The N-glycolylneuraminic acid standard is a quantitative standard of NIST-F and USP traceable Neu5Gc monosaccharide.




Figure 1: LudgerSep-R2 HPLC profile of 1,2-diamino-4,5-methylenedioxybenzene.2HCl (DMB) labeled NeuGc standard (Cat. #: CM-NEUGC-01, Batch B7A4-01).

Accuracy: This analysis was performed on 12 vials of CM-NEUGC-01.

The dispensed amount was determined as 1.04 ± 0.04 nmol NeuGc per vial.

Values are in nmols \pm standard deviation.

DMB labeled sialic acid standards eluted under the following HPLC conditions:

Column: LudgerSep R1 **(Cat. #: LS-R1-4.6x150)**

Flow: 0.5 ml/min.

Temperature: 30 °C

Solvent A: methanol:acetonitrile:water (7:9:84)

Solvent B: acetonitrile

Gradient:

Time (min)	%B
0.0	0.0
19.0	0.0
19.5	90.0
23.5	90.0
24.0	0.0
30.0	0.0
35.0	0.0

Detector: Fluorescence

Excitation wavelength: 373 nm

Emission wavelength: 448 nm