



Ludger

**Product Guide for Ludger Clean™  
Procainamide  
Clean-up Plate**

**Part of the Ludger-Velocity™ Fast Glycan Analysis Range.**

**(Ludger Product Guide: LC-PROC-96)**

**Ludger Document # LC-PROC-96-Guide-v1.0**

**Ludger Ltd**

Culham Science Centre  
Oxford OX14 3EB  
United Kingdom

Tel: +44 1865 408 554

Fax: +44 870 163 4620

Email: [info@ludger.com](mailto:info@ludger.com)

[www.ludger.com](http://www.ludger.com)

# Contents

---

Contents .....	2
Post-procainamide Labeling Clean-up Plate .....	3
Kit Contents .....	4
Additional Reagents and Equipment Required .....	4
Safety and Handling .....	4
Clean-up procedure .....	5
Time Line for Procedure .....	5
Method.....	5
Warranties and liabilities .....	7
Document Revision Number .....	7
Appendix 1 : Troubleshooting Guide .....	8

## Post-procainamide Labeling Clean-up Plate

---

<b>Applications</b>	<b>Post-procainamide Cleanup</b> For removal of excess procainamide label after procainamide labeling of glycans. This clean-up plate is used where excess procainamide interferes with subsequent sample analysis.
<b>Description</b>	The LC-PROC-96 plate is a 96 well membrane-bottom plate containing a specialized glycan binding membrane. <b>This product is designed for use with both the vacuum manifold that can be purchased from Ludger or with other popular vacuum manifold systems.</b> Excess procainamide label passes through the membrane whilst glycans are bound to the membrane allowing separation of these two components.
<b>Number of Samples</b>	Sufficient for up to 96 samples.
<b>Volume of Sample</b>	Up to 350 $\mu$ L per well.
<b>Suitable Samples</b>	Procainamide labeled glycans. Typical samples would be procainamide labelled glycans from up to 100 $\mu$ g of IgG glycoprotein.
<b>Storage</b>	Store at room temperature. Protect from sources of heat, light, and moisture. When stored correctly, the reagents should be stable for 36 months from date of purchase.
<b>Shipping</b>	The product should be shipped at ambient temperature.

**For research use only. Not for human or drug use**

## Kit Contents

---

The kit contains the following materials:

Cat. #	Item	Quantity
LC-PROC-96	LudgerClean 96 well procainamide labelled glycan clean-up plate	1

## Additional Reagents and Equipment Required

---

For a full list of vacuum manifold accessories see the Ludger-Velocity-Guide available from our website or upon request.

- Pure water: resistivity above 18 M $\Omega$ -cm, particle free (>0.22  $\mu$ m), TOC <10 ppb
- 70 % ethanol solution
- Acetonitrile (HPLC grade)
- Vacuum manifold suitable for 96 well format SPE plates (cat. no. LC-VAC-MANIFOLD Kit)
- Vacuum trap (cat. No. LC-VACUUM-TRAP-KIT)
- 2 mL collection plate for collecting glycans (cat. No. LP-COLLPLATE-2ML-96).
- Collection plate lid (optional) (cat. No. LP-COLLPLATE-LID-96).

## Safety and Handling

---

- Ensure that any glass, plasticware or solvents used with this item are free of environmental carbohydrates. Use powder-free gloves for all sample handling procedures and avoid contamination with environmental carbohydrate.
- Once used, the plate should be discarded according to local safety rules.

## Clean-up procedure

---

### Time Line for Procedure – 96 samples

Procedure	Time
Assemble the vacuum manifold	5 min
Preparation of procainamide clean-up plate	30 min
Adding sample	10 min
Eluting glycans	30 min
Drying glycans	8 hours
Total time	2 hour 15 min plus drying time

## Method

---

### 1 Assemble the vacuum manifold

Either follow your usual procedure for assembling your current vacuum manifold system or follow the instructions supplied with a Ludger vacuum manifold system (LC-VAC-MANIFOLD-KIT – see Ludger-Velocity-Guide for kit manual). Put a collection plate or other suitable container inside the manifold to collect waste (an empty pipette tip box usually fits). Place the top back on the manifold and place the procainamide clean-up plate on top of the manifold.

### 2 Preparation of the procainamide clean-up plate

Pipette 200 µL of a 70 % ethanol solution into the plate wells that are to be used to wet the membrane. Apply a vacuum and adjust to between -0.05 and -0.2 bar until all the ethanol solution has all gone through the wells. Open the tap to release the vacuum. Pipette 200 µL of water into each well to wash the membrane. Apply a vacuum and adjust to between -0.05 and -0.2 bar until the water has all gone through the wells. Pipette 200 µL of acetonitrile into each well to prime the membrane. Apply a vacuum and adjust to between -0.05 and -0.2 bar until the acetonitrile has all gone through the wells.

When applying the vacuum, you may have to push the base plate down onto the manifold until the vacuum takes hold. **The maximum pressure used should be no more than -0.5 bar**

### 3 Prepare the procainamide labelled glycan samples

Pipette 230 µL of acetonitrile into the procainamide labelled glycan samples (typically labelling mix + glycan + water is a 20 µL volume of sample). Gently mix the sample by pipette action.

#### **4 Apply the samples to the plate**

Load each sample into a primed well of the plate. Without applying a vacuum allow the sample to settle into the cartridge for 5 minutes. Apply a vacuum and adjust to between -0.05 and -0.1 bar until the acetonitrile sample solution has all gone through the wells. Discard the waste.

#### **5 Wash off non-glycan contaminants**

Add 200 µL of acetonitrile to each well containing sample. Apply a vacuum and adjust to between -0.05 and -0.1 bar until the acetonitrile has all gone through the wells. Repeat with two additional washes of 200 µL of acetonitrile. Discard the waste.

#### **6 Elute the labeled glycans**

Place a 96 well collection plate (LP-COLLPLATE-2ML) inside the vacuum manifold. Assemble the manifold with the procainamide clean-up plate on top ensuring that the collection plate is in-line with the wells. Ensure that the distance between the collection plate and the manifold top is as small as possible to reduce the gap between the procainamide clean-up plate and the collection plate (spacers may be required to lift the collection plate).

Add 100 µL of water to each well containing sample. Apply a vacuum and adjust to between -0.05 and -0.1 bar until the liquid has all gone through the wells. Open the tap to release the vacuum. Apply a higher vacuum (-0.5 bar) to remove as much remaining water from the wells (and underneath the membrane) as possible. Open the tap to release the vacuum. Repeat the higher vacuum (-0.5 bar) step for a second time.

Add another 100 µL of water into each well to wash through any remaining procainamide labelled sample. Apply a vacuum and adjust to between -0.05 and -0.1 bar until the liquid has all gone through the wells. Open the tap to release the vacuum. Apply a higher vacuum (-0.5 bar) to remove as much remaining water from the wells (and underneath the membrane) as possible. Open the tap to release the vacuum. Repeat the higher vacuum (-0.5 bar) step two additional times. These vacuum steps are done to remove as much remaining water from the wells (and underneath the membrane) as possible.

#### **7 Dry down the samples**

Samples can be stored for several hours at 4°C or for a longer period at -20°C. Samples can be stored in the collection plate, covered with a lid (LP-COLLPLATE-LID-96), or transferred to 0.5 mL centrifuge vials. Optionally, dry the samples down completely using a rotary speed vac (approximately 8 hours). We do not recommend applying heat at this stage. Only use a good quality vacuum centrifuge as long drying times or heat may lead to glycan desialylation.

## Warranties and liabilities

---

Ludger warrants that the above product conforms to the attached analytical documents. Should the product fail for reasons other than through misuse Ludger will, at its option, replace free of charge or refund the purchase price. This warranty is exclusive and Ludger makes no other warrants, expressed or implied, including any implied conditions or warranties of merchantability or fitness for any particular purpose.

Ludger shall not be liable for any incidental, consequential or contingent damages.

This product is intended for *in vitro* research only.

## Document Revision Number

---

Document # LC-PROC-96-guide, version v1.0

## Appendix 1 : Troubleshooting Guide

---

The following is a guide to the most likely problems associated with the use of the LudgerClean procainamide clean-up plate.+

### **Liquid does not flow.**

The membrane requires pre-wetting with ethanol otherwise aqueous solutions will not flow through the membrane.