



Certificate of Conformity

LudgerSep N1 Guard Column

Cat. #: LS-N1-4.6x10

Serial #: B38S-01

Size: 4.6 x 10 mm

This column conforms to the specifications given in Ludger document # LS-N1-GUIDE.

Packed Solvent 75 % acetonitrile (aq)

Test Analysis

Flow Rate 1 ml / min

Eluent Isocratic 75 % acetonitrile (aq)

Pressure 0.2 MPa

Temperature Ambient