



Ludger

Certificate of Analysis

LudgerSep N2 Amide HPLC Glycan Analysis Column

Cat. #: LS-N2-2.0x150

Serial #: B3AM-03

Size: 2.0 x150mm

This column conforms to the specifications given in Ludger document # LS-N2-GUIDE.

Packed Solvent 85 % acetonitrile (aq)

Test Analysis

Flow Rate 0.20 ml / min

Eluent Isocratic 85 % acetonitrile (aq)

Pressure 5.0 MPa


Temperature 25°C

Sample Uracil 35mg/L

Sample Volume 2µl

Detection UV 254nm

Inspection Chromatogram



Theoretical Plates = 14124

Asymmetry Factor = 0.95