



Certificate of Analysis

LudgerSep N2 Amide HPLC Glycan Analysis Column

Cat. #: LS-N2-4.6x150

Serial #: B73L-04

Size: 4.6 x150mm

This column conforms to the specifications given in Ludger document # LS-N2-GUIDE.

Packed Solvent 85 % acetonitrile (aq)

Test Analysis

Flow Rate 1.0 ml / min

Eluent Isocratic 85 % acetonitrile (aq)

Pressure 5.0 MPa


Temperature 40 °C

Sample Uracil 35mg/L

Sample Volume 10 µl

Detection UV 254nm

Inspection Chromatogram



Theoretical Plates = 19869

Asymmetry Factor = 0.98