



Certificate of Analysis

LudgerSep™ BPT Solvent

Cat. # LS-R-BPTX10

Batch #: B417-02

Size: 50 mL

Test Analysis

Column	LudgerSep uR2 LS-UR2-2.1x50			
Flow Rate	0.4 mL / min			
Solvent A	LS-R-BPTX10 made up to 500 mL with purified water			
Solvent B	Acetonitrile			
Gradient	Time (min)	% A	%B	
	00.00	0.0	0.0	
	01.00	0.0	0.0	
	04.50	84.3	15.7	
	04.60	50.0	50.0	
	05.60	50.0	50.0	
	06.00	100.0	0.0	0.6 mL/min
	08.00	100.0	0.0	0.6 mL/min

Temperature	35 °C
Sample	CM-Mono-Mix-10, B3BE-01, 1 in 50 dilution 2-AA labelled Monosaccharide Reference Standard
Injection Volume	5 µL
Fluorescence Detection	λ_{Ex} 250nm, λ_{Em} 425nm




Figure 1. LudgerSep-uR2 profile of 2AA labelled CM-Mono-Mix-10. Date of Analysis – 10 Jan 2014.