



## Certificate of Analysis

---

### LudgerSep R1 HPLC Column for Glycan Analysis

Cat. # : LS-R1-4.6x150

Batch # : B153-10

Size : 4.6 x 150 mm

---

This column conforms to the specifications given in Ludger document # LS-R1-GUIDE.

**Packed Solvent**      60 % acetonitrile (aq)

#### Test Analysis

**Flow Rate**            1.25 ml / min


**Eluent**                Isocratic 60 % acetonitrile (aq)

**Pressure**             2680 psi

**Temperature**        Ambient

**Sample Volume**     5  $\mu$ l

**Detection**            UV at 254nm



Peak No	Peak Name	Retention time (minutes)	Capacity Factor
1	Theophyline	1.18	0.00
2	p-Nitroaniline	1.72	0.46
3	Methyl Benzoate	2.59	1.21
4	Phenetole	3.63	2.09
5	o-Xylene	5.15	3.38

Efficiency o-Xylene (n/m) = 167782

Asymmetry Factor (10%) o-Xylene = 1.123