



## Certificate of Analysis

---

### LudgerSep UR2 HPLC Column for Glycan Analysis

Cat. #: LS-UR2-2.1x100

Batch #: B436-05

Size: 2.1 x 100 mm

---

This column conforms to the specifications given in Ludger document # LS-UR2-GUIDE.

**Packed Solvent** 50 % acetonitrile (aq)

#### Test Analysis

**Flow Rate** 500ul / min


**Eluent** Isocratic 50 % acetonitrile (aq)

**Pressure** 8127 psi

**Temperature** Ambient

**Sample Volume** 1  $\mu$ l

**Detection** UV at 254nm



Peak No	Peak Name	Retention time (minutes)	Tailing Factor (5%)
1	Theophylline	0.43	1.51
2	p-Nitroaniline	0.69	1.28
3	Methyl Benzoate	1.12	1.15
4	Phenetole	1.70	1.04
5	o-Xylene	2.54	1.02

Efficiency o-Xylene (n/m) = 175351

Tailing Factor (5%) o-Xylene = 1.02