

Certificate of Analysis

LudgerSep™ UR2 UHPLC Column for Glycan Analysis

Cat. #: LS-UR2-2.1x100

Batch: B9BI-01

Size: 2.1 x 100 mm

This column conforms to the specifications given in Ludger document # LS-UR2-GUIDE.

Packed Solvent 50 % acetonitrile (aq)

Test Analysis


Flow Rate 500ul / min

Eluent Isocratic 50 % acetonitrile (aq)

Pressure 7820 psi

Sample Volume 1 µl

Detection UV at 254nm



Peak No	Peak Name	Retention time (minutes)	Tailing Factor (5%)
1	Theophylline	0.49	1.34
2	p-Nitroaniline	0.78	1.25
3	Methyl Benzoate	1.27	1.20
4	Phenetole	1.93	1.11
5	o-Xylene	2.90	1.08

Efficiency o-Xylene (n/m) = 186840

Tailing Factor (5%) o-Xylene = 1.08