


Certificate of Analysis**LudgerSep™ UR2 UHPLC Column for Glycan Analysis**

Cat. #: LS-UR2-2.1x100

Batch: B9BI-04

Size: 2.1 x 100 mm

This column conforms to the specifications given in Ludger document # LS-UR2-GUIDE.

Packed Solvent 50 % acetonitrile (aq)**Test Analysis****Flow Rate** 500ul / min**Eluent** Isocratic 50 % acetonitrile (aq)**Pressure** 7828 psi**Sample Volume** 1 µl**Detection** UV at 254nm

Peak No	Peak Name	Retention time (minutes)	Tailing Factor (5%)
1	Theophylline	0.48	1.36
2	p-Nitroaniline	0.78	1.21
3	Methyl Benzoate	1.27	1.18
4	Phenetole	1.93	1.11
5	o-Xylene	2.91	1.07

Efficiency o-Xylene (n/m) = 188130

Tailing Factor (5%) o-Xylene = 1.07