

## Certificate of Analysis

### LudgerSep™ UR2 UHPLC Column for Glycan Analysis

Cat. #: LS-UR2-2.1x100

Batch: C01R-08

Size: 2.1 x 100 mm

This column conforms to the specifications given in Ludger document # LS-UR2-GUIDE.

**Packed Solvent** 50 % acetonitrile (aq)

#### Test Analysis


**Flow Rate** 500ul / min

**Eluent** Isocratic 50 % acetonitrile (aq)

**Pressure** 7229 psi

**Sample Volume** 1 µl

**Detection** UV at 254nm



<b>Peak No</b>	<b>Peak Name</b>	<b>Retention time (minutes)</b>	<b>Tailing Factor (5%)</b>
1	Theophylline	0.50	1.30
2	p-Nitroaniline	0.81	1.23
3	Methyl Benzoate	1.33	1.13
4	Phenetole	2.03	1.06
5	o-Xylene	3.05	1.04

**Efficiency o-Xylene (n/m) =195730**

**Tailing Factor (5%) o-Xylene = 1.04**