

Certificate of Analysis

LudgerZyme Ceramide Glycanase Kit

Cat. #: LZ-CER-HM-KIT	Batch #: B66S-02	Size: 1 set of enzyme per kit

This kit conforms to the specifications given in Ludger document # LZ-CER-HM-KIT-Guide-v1.0

Each kit contains the following components:

Quantity per Kit	Cat #	Batch #	Component Name
1	LZ-CER-HM-10	B66R-04	LudgerZyme Ceramide Glycanase
			(Hirudo medicinalis)
1	LZ-CER-BUFFX4	B66R-02	LudgerZyme Ceramide Glycanase
			RXN buffer
1	GLIP-GM1-01	B66R-03	GM1 glycolipid (positive control)

Expiry Date: AUG 2018



Ceramide glycanase assay: 5µL of 20µU/mL LudgerZyme Ceramide Glycanase (LZ-CER-HM-10) were incubated with 5µL of 0.33µg/µL GM1 glycolipid (GLIP-GM1-01), in a 20µL reaction containing 5µL 4X LudgerZyme Ceramide Glycanase RXN buffer (LZ-CER-BUFFX4). Reaction was incubated for 24 hours at 37°C. The released glycans were 2AB labelled (LT-KAB-A2) and S-cartridge clean-up (LC-S-A6). 2AB labelled products were analysed by HILIC-HPLC.




Figure 1: HILIC HPLC profile of two 2AB labelled GLIP-GM1-01 glycans (Batch: B66R-03) released using LZ-CER-HM-KIT (Batch: B665-02)

HPLC Running Conditions:

Column: Waters BEH Glycan column 1.7um (150mm) Flow: 0.560 mL/min

Solvent A: 50 mM ammonium formate pH 4.4 Solvent B: 100 % acetonitrile Temperature: 60°C.

<u>Gradient:</u> 0 min – 22% A, 24.8 min – 42% A, 25.8 min – 60% A, 25.9 min – 100% A, 27.9 min – 100% A , 28.4 min – 22% A, 35 min – 22%

Detector: Fluorescence Detector Excitation wavelength: 330 nm Emission wavelength: 420 n
--



No contaminating exoglycosidase or endoglycosidase activities were detected (ND) with the following substrates:

<mark>α-sialidase and endoglycosidase F3:</mark> 2AB-FA2G2S2 (CAB-A2F-01)	ND
<mark>β-galactosidase and α-fucosidase:</mark> 2AB-FA2G2 (CAB-NA2F-01)	ND
<mark>β-glucosaminidase and α-fucosidase:</mark> 2AB-FA2 (CAB-NGA2F-01)	ND
α-mannosidase and endoglycosidase F1 & F2: 2AB-mannose (CAB-MAN5-01, CAB-MAN6-01, CAB-MAN7-01, CAB-MAN8-01 and CAB-MAN9-01)	ND
α-galactosidase: 2AB B2 trisaccharide	ND